



Cornell University  
Animal Health Diagnostic Center

PO Box 5786  
Ithaca, NY 14852-5786  
t. 607 253-3900  
f. 607 253-3943  
w. diagcenter.vet.cornell.edu

January 1, 2009

To: Clients of the Animal Health Diagnostic Center (AHDC)

Subject: Testing Service Update

---

*Please share this information with your practice manager and professional staff*

## **Special Submission Requirements for Equine Infectious Anemia (EIA) Testing**

All test kits that are used for testing of horses for infection by equine infectious anemia (EIA) virus must be approved by the USDA's Center for Veterinary Biologics (CVB). In the licensing process, a manufacturer must submit test performance data to CVB as part of the approval process. Data that is submitted includes the type of samples that were used to achieve the level of test performance reported. This information defines how the test can be used. For example, the test kit information will specify whether for a serological test one can use serum or plasma. If the test kit specifies "serum", then under QA standards, the test is not valid if plasma is used. The definition of what is an "acceptable" sample is a key component of the test kit as approved.

For EIA testing, an example of an acceptable specimen as defined by several kits is: "**Specimens may be stored at 2-7C for up to five days. If longer storage is desired, store at -20C (-4F)**". The simple interpretation of this statement is that the specimen must be tested within 5 days of sampling if it is maintained at 2-7C during this period. This does not mean that samples can be shipped on day 5 and meet the kit requirements. In the past, many labs have ignored this type of restriction on the acceptability of the specimens. However, all laboratories certified by the USDA to perform EIA testing, both public and private, are required to follow kit manufacturer's instructions and the issue of the age of the specimen for EIA testing can no longer be ignored. Laboratories are responsible for enforcing the sample requirements and the failure to adhere to the kit instructions can be the basis for losing their USDA license for performing EIA tests.

A rapid education process must be instituted to inform all practitioners of these specimen requirements. "Acceptable" sample requirements have not been strictly defined in the past, but failure to adhere to the specimen requirements in the future may result in test delays, as "expired" samples will not be tested. **To ensure that testing is not delayed, the best procedure is to remove the serum from the clot tube and freeze the serum.** There is no time limit for testing frozen serum. Frozen serum should be shipped with refrigeration Overnight so the sample arrives frozen. A statement attesting to the proper handling of the samples signed by the submitting veterinarian is also advisable should the sample thaw during transit. Otherwise the laboratory has no way of knowing if the sample was handled properly when the sampling date exceeds the 5-day limit.

If there are questions as to the acceptability of a specimen, please contact the AHDC **before** submitting the specimen. Unacceptable specimens can not be returned to the submitter. In addition, EIA test paperwork for all New York horses tested at any laboratory may be reviewed by Ag & Markets staff to determine the time elapsed between when the sample was taken and when it was tested.